

- 6.9 GB/s PCIe Gen 3 (8-lane) interface
- 2 channels sampled at 12-bit resolution
- 1 GS/s real-time sampling rate
- FPGA-based FFT processing
- Variable frequency external clocking
- Continuous streaming mode
- ±400 mV fixed input range
- AlazarDSO[®] oscilloscope software
- Software Development Kit supports C/C++, C#, Python, MATLAB[®], LabVIEW[®]
- Support for Windows[®] & Linux[®]



Product	Bus	Operating System	Channels	Max. Sample Rate	Bandwidth	Memory Per Channel	Resolution
ATS9371	PCIe x8 Gen 3	32-bit/64-bit Windows & 64-bit Linux	2	1 GS/s on 2 channels	1.0 GHz	2 Gigasamples on 2 channels	12 bits

Overview

AlazarTech ATS*9371 is an 8-lane PCI Express Gen 3 (PCIe x8), dual-channel, high-speed, 12-bit, 1 GS/s waveform digitizer card capable of acquiring data into its on-board 8 GB memory or streaming acquired data to PC memory at rates up to 6.9 GB/s.

It is also possible to do FPGA-based 4096-point FFT on acquired data. This is useful for Optical Coherence Tomography (OCT) related applications.

There are two A/D converters on the ATS9371 board, each running at 1 GS/s. Unlike other products on the market, ATS9371 does not use interleaved sampling. Each input has its own 12-bit, 1 GSPS ADC chip.

Optional variable frequency external clock allows operation from 1 GHz down to 300 MHz (or 100 MHz for screened ATS9371 cards), making ATS9371 an ideal waveform digitizer for many applications.

Users can capture data from one trigger or a burst of triggers. Users can also stream very large datasets continuously to PC memory or hard disk.

ATS9371 is supplied with AlazarDSO software that lets the user start data acquisition immediately, without having to go through a software development process.


Users who need to integrate the ATS9371 in their own program can purchase a software development kit, ATS-SDK, for C/C++, C#, Python, MATLAB, and LabVIEW for both Windows and Linux operating systems.

All of this advanced functionality is packaged in a low power, half-length PCI Express Gen 3 card.

Applications

Optical Coherence Tomography (OCT)
Ultrasonic & Eddy Current NDT/NDE
RF Signal Recording & Analysis
Terabyte Storage Oscilloscope
High-Resolution Oscilloscope
Spectroscopy

Multi-Channel Transient Recording





PCI Express Gen 3 Bus Interface

ATS9371 interfaces to the host computer using an 8-lane PCI Express bus. Each lane operates at 8.0 Gbps (Gen 3).

According to PCIe specification, an 8-lane board can be plugged into any 8-lane or 16-lane slot, but not into a 4-lane or 1-lane slot. As such, ATS9371 requires at least one free 8-lane or 16-lane slot on the motherboard.

ATS9371 is fully compatible with motherboards of all generations of PCI Express (Gen 1, Gen 2 or Gen 3). At run-time, ATS9371 and the motherboard negotiate the appropriate link speed and width.

The physical and logical PCIe Gen 3 x8 interface is provided by an on-board FPGA, which also integrates acquisition control functions, memory management functions, acquisition datapath and DSP logic. This very high degree of integration maximizes product reliability.

The AlazarTech® 6.9 GB/s benchmark was done on an ASUS WS X299 SAGE motherboard.

Users must always be wary of throughput specifications from manufacturers of waveform digitizers. Some unscrupulous manufacturers tend to specify the raw, burst-mode throughput of the bus. AlazarTech, on the other hand, specifies the benchmarked sustained throughput. To achieve such high throughput, a great deal of proprietary memory management logic and kernel mode drivers have been designed.

Analog Input

An ATS9371 features two analog input channels. Each channel has up to 1.0 GHz of full power analog input bandwidth. Input voltage range is fixed at ±400 mV.


It must be noted that input impedance of both channels is fixed at 50 Ω . Input coupling is fixed to DC.

Acquisition System

ATS9371 PCI Express digitizers use state-of-the-art dual 1 GS/s, 12-bit ADCs to digitize the input signals.

The two channels are guaranteed to be simultaneous, as the two ADCs use a common clock. Note that it is not possible to perform dual edge sampling (DES) on ATS9371.

An acquisition can consist of multiple records, with each record being captured as a result of one trigger event. A record can contain both pre-trigger and post-trigger data. Up to 8064 pre-trigger points can be captured in single channel mode and 3968 in dual-channel mode. ATS9371 can capture an infinite number of triggers. In between the multiple triggers being captured, the acquisition system is re-armed by the hardware within 256 sampling clock cycles.



ATS9371 Sampling

This mode of capture, sometimes referred to as Multiple Record, is very useful for capturing data in applications with a very rapid or unpredictable trigger rate. Examples of such applications include medical imaging, ultrasonic testing, OCT and NMR spectroscopy.

On-Board Acquisition Memory

ATS9371 features two DDR3 SODIMM sockets that can each be populated with a 4 GB SODIMM, for a total on-board memory of 8 GB (4 Gigasamples).

This on-board memory is used as a very deep FIFO to temporarily store acquired ADC data before transferring it to motherboard memory using proprietary DMA engines. This on-board buffer allows loss-less data transfer even if the computer is temporarily interrupted by other tasks.

Maximum Sustained Transfer Rate

PCI Express support on different motherboards may vary, resulting in non-optimal data transfer rates. The reasons behind these differences are complex and varied and will not be discussed here.

ATS9371 users can quickly determine the maximum sustained transfer rate for their motherboard by inserting their card in a PCIe slot and running the bus benchmarking tool provided in AlazarDSO for Windows or AlazarFrontPanel for Linux.


Recommended Motherboards or PCs

Many different types of motherboards and PCs have been benchmarked by AlazarTech. The ones that have produced the best throughput results (up to 6.9 GB/s) are listed here: www.alazartech.com/images-media/2246-AlazarTechRecommendedMotherboards.pdf.

AlazarTech recommends that customers not use SandyBridge CPUs with ATS9371.

FPGA-Based FFT Processing

It is possible to do real-time FFT signal processing using the on-board FPGA. Note that only one input can be processed.



Up to 4096-point FFT length is supported. A user programmable complex windowing function can be applied to the acquired data before FFT calculation.




The complex FFT output is converted to magnitude in single-precision floating-point format. A logarithmic output is also available.

It is also possible to DMA both frequency and time domain data. This allows users to verify FPGA-based FFT operation during algorithm development.

ATS9371 can perform 250,000 4096-point FFTs per second. FPGA-based FFT is ideal for customers in the Optical Coherence Tomography (OCT) field.

FPGA-Based Digital Signal Processing

In addition to providing the bus interface and managing the acquisition engine, ATS9371's on-board FPGA is also used for digital signal processing, such as Fast Fourier Transforms. ATS9371 features a large Stratix V FPGA 5SGXMA3K3F40C3N.




ATS9371 FPGA

No Pre-Trigger (NPT) AutoDMA

Many ultrasonic scanning and medical imaging applications do not need any pre-trigger data: only post-trigger data is sufficient.

NPT AutoDMA is designed specifically for these applications. By only storing post-trigger data, the memory bandwidth is optimized and the entire onboard memory acts like a very deep FIFO.



Note that a DMA is not started until (RecordsPerBuffer +1) number of records (triggers) have been acquired and written to the on-board memory.

NPT AutoDMA buffers do not include headers. However, users can specify that each record should come with its own footer that contains a 40-bit trigger timestamp. The footer is called NPT Footer.

More importantly, a BUFFER_OVERFLOW flag is asserted only if the entire on-board memory is used up. This provides a very substantial improvement over Traditional AutoDMA.


NPT AutoDMA can easily acquire data to PC host memory at the maximum sustained transfer rate of the motherboard without causing an overflow.

This is the recommended mode of operation for most ultrasonic scanning, OCT and medical imaging applications.

It should be noted that even though this mode is called "No Pre Trigger", it is now possible to do limited pretrigger data captures, i.e. up to 8192 points in single channel mode and 4096 points in dual channel mode.

Continuous AutoDMA

Continuous AutoDMA is also known as the data streaming mode. In this mode, data starts streaming across the PCIe bus as soon as the ATS9371 is armed for acquisition. It is important to note that triggering is disabled in this mode.



Continuous AutoDMA buffers do not include headers, so it is not possible to get trigger time-stamps. A BUFFER_OVERFLOW flag is asserted only if the entire on-board memory is used up.


The amount of data to be captured is controlled by counting the number of buffers acquired. Acquisition is stopped by an AbortCapture command.

Continuous AutoDMA can easily acquire data to PC host memory at the maximum sustained transfer rate of the motherboard without causing an overflow. This is the recommended mode for very long signal recording.

Triggered Streaming AutoDMA

Triggered Streaming AutoDMA is virtually the same as Continuous mode, except the data transfer across the bus is held off until a trigger event has been detected. Triggered Streaming AutoDMA buffers do not include headers, so it is not possible to get trigger time-stamps.





A BUFFER_OVERFLOW flag is asserted only if the entire on-board memory is used up.

As in Continuous mode, the amount of data to be captured is controlled by counting the number of buffers acquired. Acquisition is stopped by an AbortCapture command.

Triggered Streaming AutoDMA can easily acquire data to PC host memory at the maximum sustained transfer rate of the motherboard without causing an overflow. This is the recommended mode for RF signal recording that has to be started at a specific time, e.g. based on a GPS pulse.

Asynchronous DMA Driver

The various AutoDMA schemes discussed above provide hardware support for optimal data transfer. However, a corresponding high-performance software mechanism is also required to make sure sustained data transfer can be achieved.

This proprietary software mechanism is called Async DMA (short for Asynchronous DMA).

A number of data buffers are posted by the application software. Once a data buffer is filled, i.e. a DMA has been completed, ATS9371 hardware generates an interrupt, causing an event message to be sent to the application so it can start consuming data. Once the data has been consumed, the application can post the data buffer back on the queue. This can go on indefinitely.

One of the great advantages of Async DMA is that almost 95% of CPU cycles are available for data processing, as all DMA arming is done on an event-driven basis.

To the best of our knowledge, no other supplier of waveform digitizers provides asynchronous software drivers. Their synchronous drivers force the CPU to manage data acquisition, thereby slowing down the overall data acquisition process.

Data Packing Mode

By default, ATS9371 stores 12-bit data acquired by its on-board A/D converters as a 16-bit integer. Users can also choose to pack the data as 12-bit integers or even 8-bit integers. Being able to reduce the total amount of data being transferred can be very useful in data recording applications.

Note that it is the user application's responsibility to unpack the data. Also note that NPT Footers are not available in Data Packing Mode.

Triggering

ATS9371 is equipped with sophisticated digital triggering options, such as programmable trigger thresholds and slope on any of the input channels or the External Trigger input.

While most oscilloscopes offer only one trigger engine, ATS9371 offers two trigger engines (called Engines J and K).

The user can specify the number of records to capture in an acquisition, the length of each record and the amount of pre-trigger data.

A programmable trigger delay can also be set by the user. This is very useful for capturing the signal of interest in a pulse-echo application, such as ultrasound, radar, lidar etc.

External Trigger Input

ATS9371 external trigger input (TRIG IN) can be set as an analog input with ± 2.5 V full scale input range and 50 Ω input impedance, or a 3.3 V TTL input.

When TTL input is selected, the input impedance increases to approximately 6.6 $k\Omega,$ making it easier to drive the TRIG IN input from high-output impedance sources.

Note: If full 12-bit resolution is required, users should select CH A or CH B as the trigger source. When the External Trigger Input is used as the trigger source, the least significant bit (LSB) of each 12-bit sample is replaced by the state of the external trigger signal source.

Timebase

ATS9371 timebase can be controlled either by onboard low-jitter VCO or by optional External Clock.

On-board low-jitter VCO uses a 10 MHz TCXO as a reference clock. Clock buffers used feature less than 76 fs $_{\!\!\!\!\!\!\!RMS}$ additive jitter.

Optional External Clock

While the ATS9371 features low jitter VCO and a 10 MHz TCXO as the source of the timebase system, there may be occasions when digitizing has to be synchronized to an external clock source.

ATS9371 External Clock option provides an SMA input for an external clock signal, which should have a high slew rate. Signal levels, specified in detail on page 8, must be respected.

Input impedance for the External Clock input is fixed at 50 Ω . External clock input is always AC-coupled.

There are two types of External Clock supported by ATS9371: Fast External Clock and 10 MHz Reference.



Fast External Clock

A new sample is taken by the on-board ADCs for each rising edge of this External Clock signal.

In order to satisfy the clocking requirements of the ADC chips being used, Fast External Clock frequency must always be higher than 300 MHz and lower than 1 GHz.

For customers whose external clocks may go lower than 300 MHz during the acquisition, it is possible to have AlazarTech screen the ATS9371 boards for external clock operation down to 100 MHz (Order number ATS9371-006)

This is the ideal clocking scheme for OCT applications.

10 MHz Reference Clock


It is possible to generate the sampling clock based on an external 10 MHz reference input. This is useful for RF systems that use a common 10 MHz reference clock.

ATS9371 uses an on-board low-jitter PLL to generate a user-specified high-frequency clock used by the ADC. This sampling clock can be any multiple of 1 MHz between 300 MHz and 1 GHz.

OCT Ignore Bad Clock

The ADCs used on the ATS9371 require the external clock frequency to be above 300 MHz and lower than 1 GHz. In OCT applications, these limits cannot always be respected due to the nature of the optical source.

AlazarTech's OCT Ignore Bad Clock technology allows safe operation with these out-of-specification clocks without requiring the use of a dummy clock in the source.



Firmware version 26.04+, driver version 6.00.01+ and SDK 7.1.3+ are required to take advantage of OCT Ignore Bad Clock. For existing customers, these firmware and driver versions are available for download from AlazarTech's website free of charge.

See www.alazartech.com/en/technology/oct-ignore-bad-clock/ for more information on this technology.

AUX Connector

ATS9371 provides an AUX (Auxiliary) SMA connector that is configured as a Trigger Output connector by default.

When configured as a Trigger Output, AUX SMA connector outputs a 5 Volt TTL signal synchronous to the ATS9371 Trigger signal, allowing users to synchronize their test systems to the ATS9371 Trigger.

When combined with the Trigger Delay feature of the ATS9371, this option is ideal for ultrasonic and other pulse-echo imaging applications.

AUX connector can also be used as a Trigger Enable Input for Frame Capture (B-scan) applications. In fact, this is the most popular use of AUX connector in OCT applications.

Calibration

Every ATS9371 digitizer is factory calibrated to NIST- or CNRC-traceable standards. To recalibrate an ATS9371, the digitizer must be shipped back to the factory.

On-Board Monitoring

Adding to the reliability offered by ATS9371 are the on-board diagnostic circuits that constantly monitor over 20 different voltages, currents and temperatures. LED alarms are activated if any of the values surpass the limits.

AlazarDSO Software

ATS9371 is supplied with the powerful AlazarDSO software that allows the user to setup the acquisition hardware and capture, display and archive the signals.

The Stream-To-Memory command in AlazarDSO allows users to stream a large dataset to motherboard memory.

AlazarDSO software also includes powerful tools for benchmarking the computer bus and disk drive.

Software Development Kits

AlazarTech provides easy to use software development kits for customers who want to integrate the ATS9371 into their own software.

A Windows and Linux compatible software development kit, called ATS-SDK, includes headers, libraries and source code sample programs written in C/C++, C#, Python, MATLAB, and LabVIEW. These programs can fully control the ATS9371 and acquire data in user buffers.

The purchase of an ATS-SDK license includes a subscription that provides the following benefits for a period of 12 months from the date of purchase:

- Download ATS-SDK updates from the AlazarTech website;
- Receive technical support on ATS-SDK.



Customers who want to receive technical support and download new releases beyond this 12 month period should purchase extended support and maintenance (order number ATS-SDK-1YR).

ATS-GPU

ATS-GPU is a software library developed by AlazarTech to allow users to do real-time data transfer from ATS9371 to a GPU card at rates up to 6.9 GB/s.

Interfacing waveform digitizers to GPUs involves creating a software mechanism to move data from one to the other and back to user buffers. The standard techniques used most often can get the job done, but feature very low data throughput due to software overheads.

AlazarTech designed ATS-GPU to eliminate this software bottleneck so that data can be moved from AlazarTech digitizers to GPUs and from GPUs to user buffers at full PCIe bus speeds. Once the data is available in GPU memory, many types of digital signal processing (DSP) can be done on this data at near-hardware speeds.

ATS-GPU-BASE is supplied with an example user application in source code. The application includes GPU kernels that use ATS-GPU to receive data, do very simple signal processing (data inversion), and copy the processed (inverted) data back to a user buffer. All this is done at the highest possible data transfer rate.

Programmers can replace the data inversion code with application-specific signal processing kernels to develop custom applications.

ATS-GPU-OCT is the optional OCT Signal Processing library for ATS-GPU. It contains floating-point FFT routines that have also been optimized to provide the maximum number of FFTs per second. Kernel code running on the GPU can do zero-padding, apply a windowing function, do a floating-point FFT, calculate the amplitude and convert the result to a log scale. It is also possible to output phase information.

FFTs can be done on triggered data or on continuous gapless stream of data. It is also possible to do spectral averaging. Our benchmarks showed that it was possible to do 970,000 FFTs per second when capturing data in dual-channel mode and using a NVIDIA® Ouadro® P5000 GPU.

ATS-GPU-NUFFT is an extension of ATS-GPU-OCT that allows non-uniform FFTs to be performed on data acquired uniformly in time domain using a fixed sampling rate. For SS-OCTs where the wave-length does not vary linearly in time, a fixed sampling rate results in data that is non-uniformly dis-tributed in frequency domain. ATS-GPU-NUFFT allows linearized FFTs to be performed on such data.

ATS-GPU supports 64-bit Windows and 64-bit Linux for CUDA®-based development.

Support for Windows

Windows support for ATS9371 includes Windows 10, Windows Server® 2019, and Windows Server 2016. As Windows Server 2019 and 2016 are seldom used by our customers, they are expected to work but are not regularly tested with each software release. If there are issues related to Windows Server 2016 or 2019, tech support may not be as rapid as for other operating systems.

Microsoft mainstream support ended in 2018 for Windows 8.1 and Windows Server 2012 R2. As such, AlazarTech has ceased development on these operating systems. Current software and driver releases may work with these operating systems but they are not officially supported.

Due to lack of demand and due to the fact that Microsoft no longer supports these operating systems, AlazarTech no longer supports Windows 8, Windows 7, Windows XP, Windows Vista, Windows Server 2012, Windows Server 2008 R2, and Windows Server 2008.

Linux Support

AlazarTech offers Dynamic Kernel Module Support (DKMS) drivers for the following Linux distributions: Ubuntu, Debian, and RHEL®.

AlazarTech DKMS drivers may work for other Linux distributions but they have not been tested and technical support may be limited.

Users can download the DKMS driver for their specific distribution by choosing from the available drivers here:

ftp://release@ftp.alazartech.com/outgoing/linux

Only 64-bit Linux operating systems are supported.

A GUI application called AlazarFrontPanel that allows simple data acquisition and display is also provided.

ATS-SDK includes source code example programs for Linux, which demonstrate how to acquire data programmatically using a C compiler.

Based on a minimum annual business commitment, the Linux driver source code license (order number ATS9371-LINUX) may be granted to qualified OEM customers for a fee. For release of driver source code, a Non-Disclosure Agreement must be executed between the customer's organization and AlazarTech.

All such source code disclosures are made on an as-is basis with limited support from the factory.

Extended Warranty

The purchase of an ATS9371 includes a standard one (1) year parts and labor warranty. Customers may extend their warranty by ordering an Extended Warranty (order number ATS9371-061).



This must be purchased before expiration of the standard warranty (or before expiration of an Extended Warranty). Extended Warranties can only be purchased while there is a valid warranty in place.

Users can purchase up to 4 (four) additional years of warranty extensions for a maximum total of 5 years of warranty.

Get your warranty end date by registering your product at: www.alazartech.com/en/my-account/my-products/.

Export Control Classification

According to the Export Controls Division of the Government of Canada, ATS9371 is currently not controlled for export from Canada. Its export control classification is N8, which is equivalent to ECCN EAR99. ATS9371 can be shipped freely outside of Canada, with the exception of countries listed on the Area Control List and Sanctions List. Furthermore, if the end-use of ATS9371, in part or in its entirety, is related to the development or deployment of weapons of mass destruction, AlazarTech is obliged to apply for an export permit.

RoHS Compliance

ATS9371 is fully RoHS compliant, as defined by Directive 2015/863/EU (RoHS 3) of the European Parliament and of the Council of 31 March 2015 on the restriction of the use of certain hazardous substances in electrical and electronic equipment.

All manufacturing is done using RoHS-compliant components and lead-free soldering.

EC Conformity

ATS9371 conforms to the following standards:

Electromagnetic Emissions:

CISPR 32:2015/AMD1:2019 /

EN 55032:2015/A11:2020 (Class A):

Multimedia Equipment (MME) Radio disturbance characteristics. Limits and method of measurement: EN 61000-3-2:2014, EN 61000-3-3:2013.

Electromagnetic Immunity:

EN 55035:2017/A11:2020:

Multimedia Equipment (MME) Immunity characteristics — Limits and methods of measurement: EN 61000-4-2:2009.

Safety:

IEC 62368-1:2014 / EN 62368-1:2014+A11:2017: Audio/video, information and communication technology equipment - Part 1: Safety requirements.

ATS9371 also follows the provisions of the following directives: 2014/35/EU (Low Voltage Equipment); 2014/30/EU (Electromagnetic Compatibility).

FCC & ICES-003 Compliance

ATS9371 has been tested and found to comply with the limits for a Class A digital device, pursuant to part 15, subpart B of the FCC Rules, and the Canadian Interference-Causing Equipment Standard ICES-003 issue 7 October 2020.



System Requirements

Personal computer with at least one free x8 or x16 PCI Express slot (must be Gen 3 slot to achieve full data throughput), 16 GB RAM, 100 MB of free hard disk space, SVGA display adaptor and monitor with at least a 1024 x 768 resolution.

Power Requirements

+12 V 1.5 A, typical +3.3 V 3.0 A, typical

Physical

Size Single slot, half length PCI Express

card (4.377 inches x 6.5 inches excluding the connectors protruding from the front panel)

Weight 250 g

I/O Connectors

ECLK, CH A, CH B,
TRIG IN, AUX I/O SMA female connector

Environmental

Operating temperature 0 to 55 degrees Celsius
Storage temperature -20 to 70 degrees Celsius
Relative humidity 5 to 95%, non-condensing

Acquisition System

Resolution 12 bits

Bandwidth (-3 dB)

DC-coupled, $50~\Omega$ Standard DC - 1.0 GHz Number of channels 2, simultaneously sampled

Maximum Sample Rate 1 GS/s single shot

Minimum Sample Rate 1 KS/s single shot for internal

clocking

Full Scale Input ranges

50 Ω input impedance: ±400 mV

DC accuracy $\pm 2\%$ of full scale in all ranges

Input coupling DC Input impedance 50 Ω ±1%

Input protection

50 Ω

±4 V (DC + peak AC for CH A, CH B and TRIG IN only without

external attenuation)

Acquisition Memory System

Memory Size 8 GB (4 Gigasamples in one

channel mode)

Record Length Software-selectable with 128-point

resolution. Record length must be a minimum of 256 points. There is no upper limit on the maximum

record length.

Number of Records Software selectable from a

minimum of 1 to a maximum of

infinite number of records

Pre-trigger depth From 0 to 8176 for single channel

From 0 to 4088 for dual channel

Post-trigger depth Record Length - Pre-Trigger Depth

Timebase System

Timebase options Internal Clock or

External Clock (Optional)

Internal Clock accuracy ±2 ppm

Internal Sample Rates 1 GS/s, 800 MS/s, 500 MS/s, 200 MS/s, 100 MS/s, 500 MS/s

200 MS/s, 100 MS/s, 50 MS/s, 20 MS/s, 10 MS/s, 5 MS/s, 2 MS/s, 1 MS/s, 500 KS/s, 200 KS/s, 100 KS/s, 50 KS/s, 20 KS/s, 10 KS/s, 5 KS/s, 2 KS/s, 1 KS/s

Dynamic Parameters

Typical values measured on the 400 mV range of CH A of a randomly selected ATS9371. Input signal was provided by an SRS SG384 signal generator, followed by a 9-pole, 100 MHz band-pass filter (TTE Q36T-100M-10M-50-720BMF). Input frequency was set at 99.9 MHz and output amplitude was set to approximately 95% of the full scale input.

SNR 57.1 dB SINAD 56.6 dB

Note that these dynamic parameters may vary from one unit to another, with input frequency and with the full scale input range selected.

Optional ECLK (External Clock) Input

Signal Level 400 mV_{P-P} to 2 V_{P-P}

 $\begin{array}{ll} \text{Input impedance} & 50 \ \Omega \\ \text{Input coupling} & \text{AC} \end{array}$

Maximum frequency

for Fast External Clock 1 GHz

Minimum frequency

for Fast External Clock 300 MHz

100 MHz for Screened ECLK boards

Sampling Edge Rising only

Optional 10 MHz Reference PLL Input

Signal Level 400 mV_{P-P} to 2 V_{P-P}

 $\begin{array}{ll} \text{Input impedance} & 50 \ \Omega \\ \\ \text{Input Coupling} & \text{AC coupled} \\ \\ \text{Input Frequency} & 10 \ \text{MHz} \pm 0.1 \ \text{MHz} \end{array}$

Maximum frequency 10.1 MHz
Minimum frequency 9.9 MHz

Sampling Clock Freq. Any multiple of 1 MHz between:

300 MHz and 1 GHz

Triggering System

Mode Edge triggering with hysteresis

Comparator Type Digital comparators for internal (CH A, CH B) triggering and

(CH A, CH B) triggering and analog comparators for TRIG IN (External) triggering

Number of Trigger Engines 2

Trigger Engine Combination Engine J, engine K, J OR K,

software selectable

Trigger Engine Source CH A, CH B, EXT, Software or None,

independently software selectable for each of the two Trigger Engines

Hysteresis $\pm 5\%$ of full scale input, typical



±10% of full scale input range. Trigger sensitivity

This implies that the trigger system may not trigger reliably if the input has an amplitude less than ±10%

of full scale input range selected

±5%, typical, of full scale input range of the selected trigger

source

Bandwidth 250 MHz

Trigger level accuracy

Trigger Delay Software selectable from 0 to

9,999,999 sampling clock cycles

Trigger Timeout Software selectable with a 10 μs

resolution. Maximum settable value is 3,600 seconds. Can also be disabled to wait indefinitely for

a trigger event

TRIG IN (External Trigger) Input

Input type Analog or 3.3 V TTL,

software-selectable

Input coupling DC only Analog input impedance 50 Ω

Analog bandwidth (-3 dB) DC - 250 MHz Analog input range ±2.5 V

Analog DC accuracy ±10% of full scale input Analog input protection

±8 V (DC + peak AC without external attenuation)

TTL input impedance $6.6 \text{ k}\Omega \pm 10\%$

TTL min. pulse width 32 ADC sampling clocks

TTL min. pulse amplitude 2 Volts

TTL input protection -0.7 V to + 5.5 V

Auxiliary I/O (AUX I/O)

Signal direction Input or Output, software select-

able. Trigger Output by default

Output types: Trigger Output,

Pacer (programmable clock) Output,

Software-controlled Digital Output

Trigger Enable Input types:

Software-readable Digital Input

Output

5 Volt TTL Amplitude:

Synchronization: Synchronized to a clock derived

from the ADC sampling clock. Divide-by-4 clock (dual channel mode) or divide-by-8 clock (single

channel mode)

Input

Amplitude: 3.3 Volt TTL (5 Volt-compliant)

DC Input coupling:

Materials Supplied

ATS9371 PCI Express Card

ATS9371 Install Disk on USB flash drive

Certification and Compliances

RoHS 3 (Directive 2015/863/EU) Compliance

CE Marking — EC Conformity

FCC Part 15 Class A / ICES-003 Class A Compliance

ORDERING INFORMATION

ATS9371 ATS9371-001

ATS9371: External Clock Upgrade ATS9371-005

ATS9371: Screened External Clock Upgrade ATS9371-006

ATS9371: One Year Extended Warranty ATS9371-061

ATS-SDK Software Development Kit

License + 1 Year Subscription

(Supports C/C++, Python, MATLAB, and LabVIEW)

ATS-GPU-BASE: GPU Streaming Library ATSGPU-001

License + 1 Year Subscription

ATS-GPU-OCT: Signal Processing Library ATSGPU-101

License + 1 Year Subscription (requires ATSGPU-001)

ATS-GPU-NUFFT: ATS-GPU-OCT Extension ATSGPU-201

for fixed-frequency sampled data License + 1 Year Subscription (requires ATSGPU-001 & ATSGPU-101)

All specifications are subject to change without notice

† AlazarDSO, AlazarTech, and AlazarTech ATS are registered trademarks of Alazar Technologies Inc.

MATLAB is a trademark and/or registered trademark of The MathWorks, Inc. LabVIEW is a trademark and/or registered trademark of National Instruments. Windows and Windows Server are trademarks and/or registered trademarks of Microsoft Corporation in the U.S. and/or other countries.

Linux is a registered trademark of Linus Torvalds.

ASUS is either a US registered trademark or trademark of ASUSTEK Computer Inc. in the United States and/or other countries.

RHEL is a registered trademark of Red Hat, Inc. in the United States and other countries.

CUDA, NVIDIA, and Quadro are trademarks and/or registered trademarks of NVIDIA Corporation in the U.S. and/or other countries.

All other trademarks are the property of their respective owners.

Manufactured By:

Alazar Technologies, Inc.

6600 TRANS-CANADA HIGHWAY, SUITE 310 POINTE-CLAIRE, QC, CANADA H9R 4S2

TOLL FREE: 1-877-7-ALAZAR OR 1-877-725-2927 TEL: (514) 426-4899 FAX: (514) 426-2723

E-MAIL: sales@alazartech.com



DATASHEET REVISION HISTORY

Updated terminology to match the standard: changed Information Technology Equipment (ITE) to Multimedia Equipment (MME) Specified Electromagnetic Immunity limits and methods of measurement: EN 61000-4-2:2009 Changes from version 1.6H (June 2021) to version 1.6I Section, Page Updated Linux Support: only 64-bit Linux operating systems are supported Updated Linux Support: only 64-bit Linux operating systems are supported Updated Linux Support: only 64-bit Linux operating systems are supported Updated Linux Support: only 64-bit Linux operating systems are supported Updated Linux Support: only 64-bit Linux operating systems are supported Updated Linux Support: only 64-bit Linux operating systems are supported Updated Linux Support: only 65-bit Linux operating systems are supported Updated Linux Support: only 65-bit Linux operating systems are supported Updated throughput from 6.8 GB/s to 6.9 GB/s Updated motherboard used for benchmarking PCI Express Gen 3 Bus Interface, pp. 2 Updated on Ars-GPU and added paragraph on ATS-GPU-NUFFT ATS-GPU, pp. 6 Updated Section ATS-GPU and added paragraph on ATS-GPU-NUFFT ATS-GPU, pp. 6 Updated product registration URL Updated section ATS-GPU and added new DKMS drives Updated product registration URL Updated standards and directives Extended Warranty, pp. 7 FCC & ICES-003 Standard FCC onformity, pp. 7 Updated standards and directives Extended Warranty Updated Sampling Rate column to Max. Sample Rate Corrected TRIG IN Input type, removed: (5 V compliant) Changed Sampling Rate column to Max. Sample Rate Replaced signal levels drive for Fast Extendel Clock and 10 MHz Reference Clock, replaced min. and max. amplitude with a note that signal levels specified on page 8 must be respected. Changed Signal level from ±200 mV to 400 mV _{P,P} to 3400 mV _{P,P} to 2 V _{P,P} " Changes from version 1.5E (Jan 2019) to version 1.5F Removed and Signal level from ±200 mV to 400 mV _{P,P} to 3400 mV _{P,P}	Changes from version 1.6I (Sept 2021) to version 1.6J	Section, Page
changed Information Technology Equipment (ITE) to Multimedia Equipment (MME) Specified Electromagnetic Immunity limits and methods of measurement: EM 61000-4-2:2009 Changes from version 1.6H (June 2021) to version 1.6I Updated support status for Windows 8.x and Windows Server versions 2012 R2, 2016, 2019 Support for Windows, pg. 6 Updated Linux Support: only 64-bit Linux operating systems are supported Updated Linux Support: only 64-bit Linux operating systems are supported Updated Linux Support: only 64-bit Linux operating systems are supported Updated Linux Support: only 64-bit Linux operating systems are supported Updated Linux Support: only 64-bit Linux operating systems are supported Updated Electromagnetic Immunity standard Changes from version 1.6G (Jan 2020) to version 1.6H Updated data throughput from 6.8 GB/s to 6.9 GB/s Updated motherboard used for benchmarking PCI Express Gen 3 Bus Interface, pg. 2 Removed 5 V-compliant from 3.3 VTTL input External Tringer Input, pg. 4 Updated section ATS-GPU and added paragraph on ATS-GPU-NUFFT Updated section ATS-GPU and added paragraph on ATS-GPU-NUFFT Updated standards and directives Updated standards and directives Updated standards and directives Updated standards and directives Updated year of ICES-003 standard FCC & ICES-003 Compliance, pg. 7 Corrected TRIG IN Input type, removed: (5 V compliant) TRIG IN (External Tringer) Input, pg. 9 Added Auxillary I/O input coupling (DC) Updated software descriptions and added order number for ATS-GPU-NUFFT Changes from version 1.5F (May 2019) to version 1.5G Changed Signal sine or square wave requirement with high slew rate Removed ATS-GPU and Support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Mindows 7 version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Mindows 7 version support, re-ordered list of operating systems, and adde	Specified number of extended warranties that users may purchase	Extended Warranty, pg. 7
Updated Linux Support: only 64-bit Linux operating systems are supported Updated Electromagnetic Immunity standard EC Conformity, pg. 6 Updated Electromagnetic Immunity standard EC Conformity, pg. 7 Changes from version 1.66 (Jan 2020) to version 1.6H Updated data throughput from 6.8 GB/s to 6.9 GB/s Updated motherboard used for benchmarking Updated motherboard used for benchmarking Removed 5 V-compliant from 3.3 V TIL input Updated oCT Ignore Bad Clock URL Updated CT Ignore Bad Clock URL Updated Linux Support (RHEL) and added new DKMS drivers Updated Linux Support (RHEL) and added new DKMS drivers Updated year of ICES-003 standard Updated year of ICES-003 standard Updated year of ICES-003 standard FCC & ICES-003 compliance, pg. 7 Updated software descriptions and added order number for ATS-GPU-NUFFT Updated software descriptions and added order number for ATS-GPU-NUFFT Changes from version 1.5F (May 2019) to version 1.5G Replaced signal sine or square wave requirement with high slew rate min. and max. amplitude with a note that signal levels pecified on page 8 must be respected. Removed qualified metrology lab as option for recalibrating ATS9371 Specified Unindows 7 version support, re-ordered list of operating systems, and added end-support notice for Windows 7 and Windows Server 2008 R2 Specified Uninva distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mVp.p" to "40	changed Information Technology Equipment (ITE) to Multimedia Equipment	t (MME)
Updated Linux Support: only 64-bit Linux operating systems are supported Updated Electromagnetic Immunity standard Changes from version 1.66 (Jan 2020) to version 1.6H Updated data throughput from 6.8 GB/s Updated data throughput from 6.8 GB/s to 6.9 GB/s Updated motherboard used for benchmarking Updated OCT Ignore Bad Clock URL Updated OCT Ignore Bad Clock URL Updated OCT Ignore Bad Clock URL Updated product registration URL Updated scandards and directives Updated scandards and updated and updated scandards and updated and updated and updated and updated and updated scandards and updated and updated and updated and updated scandards and updated and updated and updated and updated and updated u	Changes from version 1.6H (June 2021) to version 1.6I	Section, Page
Changes from version 1.6G (Jan 2020) to version 1.6H Changes from version 1.6G (Jan 2020) to version 1.6H Section, Page Updated data throughput from 6.8 GB/s to 6.9 GB/s Updated motherboard used for benchmarking Removed 5 V-compliant from 3.3 V TTL input Updated OCT Ignore Bad Clock URL Updated Section ATS-GPU and added paragraph on ATS-GPU-NUFFT Updated Linux Support (RHEL) and added new DKMS drivers Updated standards and directives Updated standards and directives Updated year of ICES-003 standard Updated section IN Input type, removed: (5 V compliant) Added Auxiliary I/O input coupling (DC) Updated software descriptions and added order number for ATS-GPU-NUFFT Changes from version 1.5F (May 2019) to version 1.5G Replaced Signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clock, replaced min. and max. amplitude with a note that signal levels specified on page 8 must be respected. Removed qualified metrology lab as option for recalibrating ATS9371 Changed signal level from "400 mV _{p-p} " to "400 mV _{p-p} to 2 V _{p-p} " Removed addinged level from 4200 mV to 400 mV _{p-p} to 2 V _{p-p} " Removed addinged level from 4200 mV to 400 mV _{p-p} to 100 mV _{p-p} to	Updated support status for Windows 8.x and Windows Server versions 2012 R2,	2016, 2019 Support for Windows, pg. 6
Changes from version 1.6G (Jan 2020) to version 1.6H Updated data throughput from 6.8 GB/s to 6.9 GB/s Updated motherboard used for benchmarking Updated motherboard used for benchmarking Removed 5 V-compliant from 3.3 V TTL input External Trigger Input, pg. 4 Updated OCT Ignore Bad Clock URL Updated Section ATS-GPU and added paragraph on ATS-GPU-NUFFT Updated Linux Support (RHEL) and added new DKMS drivers Updated standards and directives Updated standards and directives Updated standards and directives Updated year of ICES-003 standard Updated year of ICES-003 type year year year year year year year yea	Updated Linux Support: only 64-bit Linux operating systems are supported	Linux Support, pg. 6
Updated data throughput from 6.8 GB/s to 6.9 GB/s Updated motherboard used for benchmarking PCI Express Gen 3 Bus Interface, pg. 2 Removed 5 V-compliant from 3.3 V TTL input Updated OCT Ignore Bad Clock, pg. 5 Updated Section ATS-GPU and added paragraph on ATS-GPU-NUFFT Updated Linux Support (RHEL) and added new DKMS drivers Updated product registration URL Updated product registration URL Updated year of ICES-003 standard FCC & ICES-003 Standard FCC & ICES-003 Standard FCC & ICES-003 Standard FCC & ICES-003 Compliance, pg. 7 Updated year of ICES-003 standard FCC & ICES-003 Compliance, pg. 7 Updated software descriptions and added order number for ATS-GPU-NUFFT Changes from version 1.5F (May 2019) to version 1.5G Replaced signal sine or square wave requirement with high slew rate Because signal level fifter for Fast External Clock and 10 MHz Reference Clock, replaced min. and max. amplitude with a note that signal levels specified on page 8 must be respected. Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows 5 reversion 400 mVp.p.* to 400 mVp.p.p to 2 Vp.p.* Removed qualified metrology lab as option for recalibrating ATS9371 Changed signal level from *400 mVp.p.* to *400 mVp.p.p to 2 Vp.p.* Removed maximum amplitude, information is included in signal level Changed signal level from *400 mVp.p.* to *400 mVp.p.p to 2 Vp.p.* Removed darish distributions: CentOS, Debian, and Ubnutu Changed signal level from *400 mVp.p.* to *400 mVp.p.p to 2 Vp.p.* Removed darish data through a scenario sincluded in signal level Changed signal level from *400 mVp.p.* to *400 mVp.p.p to 2 Vp.p.* Removed darish data through a scenario sincluded in signal level Changed signal level from *400 mVp.p.* to *400 mVp.p.p to 2 Vp.p.* Removed darish data through a scenario sincluded in signal level Changed signal level from *400 mVp.p.p to *400 mVp.p.p to *400 mVp	Updated Electromagnetic Immunity standard	EC Conformity, pg. 7
Nemoved 5 V-compliant from 3.3 V TIL input Removed 5 V-compliant from 3.3 V TIL input Removed 5 V-compliant from 3.3 V TIL input Removed 6 V-compliant form 3.5 V TIL input Removed 6 V-compliant form 3.5 V TIL input Removed 1 Ric in input fyee, removed: (5 V compliant) Removed 6 Auxiliary I/O input coupling (DC) Removed 6 Section Past Removed 6 Signal sine or square wave requirement with high slew rate Replaced 8 Signal sine or square wave requirement with high slew rate Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from \$400 mV_{P,p}\$* to \$	Changes from version 1.6G (Jan 2020) to version 1.6H	Section, Page
Removed 5 V-compliant from 3.3 V TTL input Updated OCT Ignore Bad Clock URL Updated Section ATS-GPU and added paragraph on ATS-GPU-NUFFT Updated Linux Support (RHEL) and added new DKMS drivers Updated Jinux Support (RHEL) and added new DKMS drivers Updated sextion ATS-GPU and added new DKMS drivers Updated standards and directives Updated standards and directives Updated year of ICES-003 standard FCC & ICES-003 Compliance, pg. 7 Corrected TRIG IN Input type, removed: (5 V compliant) Added Auxiliary I/O input coupling (DC) Updated storkware descriptions and added order number for ATS-GPU-NUFFT Ordering Information, pg. 9 Changes from version 1.5F (May 2019) to version 1.5G Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clock, replaced min. and max. amplitude with a note that signal levels specified on page 8 must be respected. Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from *400 mVp_p* to *400 mVp_p* to *2 Vp_p*' Removed qualified metrology lab so solvino for recalibrating ATS9371 Changed signal level from \$400 mVp_p* to *400 mVp_p* to *2 Vp_p*' Removed in signal level from \$400 mVp_p* to *400 mVp_p* to *2 Vp_p*' Removed in signal level from \$400 mVp_p* to *400 mVp* to *	Updated data throughput from 6.8 GB/s to 6.9 GB/s	Global change
Updated OCT Ignore Bad Clock URL Updated section ATS-GPU and added paragraph on ATS-GPU-NUFFT Updated Linux Support (RHEL) and added new DKMS drivers Updated product registration URL Updated product registration URL Extended Warranty, pg. 7 Updated standards and directives EC Conformity, pg. 7 Updated standards and directives EC Conformity, pg. 7 Updated year of ICES-003 standard FCC & ICES-003 compliance, pg. 7 Corrected TRIG IN Input type, removed: (5 V compliant) TRIG IN (External Trigger) Input, pg. 9 Updated software descriptions and added order number for ATS-GPU-NUFFT Ordering Information, pg. 9 Updated software descriptions and added order number for ATS-GPU-NUFFT Ordering Information, pg. 9 Updated software descriptions and added order number for ATS-GPU-NUFFT Ordering Information, pg. 9 Updated software descriptions and added order number for ATS-GPU-NUFFT Ordering Information, pg. 9 Updated software descriptions and added order number for ATS-GPU-NUFFT Ordering Information, pg. 9 Updated software descriptions and added order number for ATS-GPU-NUFFT Ordering Information, pg. 9 Updated software descriptions and added order number for ATS-GPU-NUFFT Ordering Information, pg. 9 Updated software descriptions and added order number for ATS-GPU-NUFFT Ordering Information, pg. 9 Updated software descriptions and added order number for ATS-GPU-NUFFT Ordering Information, pg. 5 Section, Page Changes from version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Linux Support, pg. 6 Changed signal level from *400 mV _{P-P} * to *400 mV _{P-P} , to *400 mV _{P-P} , to *400 mV _{P-P} optional ECLK (External Clock) Input, pg. 8 Removed ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) Auxiliary I/O (AUX I/O), pg. 9 Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) ATS-GPU, pg. 6 Added section Extended Warranty Removed ATS-GMA soction as this product is bei	Updated motherboard used for benchmarking	PCI Express Gen 3 Bus Interface, pg. 2
Updated section ATS-GPU and added paragraph on ATS-GPU-NUFFT Updated Linux Support (RHEL) and added new DKMS drivers Linux Support, pg. 6 Updated product registration URL Extended Warranty, pg. 7 Updated standards and directives EC Conformity, pg. 7 Updated year of ICES-003 standard FCC & ICES-003 compliance, pg. 7 Corrected TRIG IN Input type, removed: (5 V compliant) Added Auxiliary I/O input coupling (DC) Updated software descriptions and added order number for ATS-GPU-NUFFT Ordering Information, pg. 9 Changes from version 1.5F (May 2019) to version 1.5G Section, Page Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clock, replaced min. and max. amplitude with a note that signal levels specified on page 8 must be respected. Removed qualified metrology lab as option for recalibrating ATS9371 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P,P} " to "400 mV _{P,P} to "400 mV _{P,P} to "400 mV _{P,P} to "400 mV _{P,P} to Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P,P} to "400 mV _{P,P}	Removed 5 V-compliant from 3.3 V TTL input	External Trigger Input, pg. 4
Updated Linux Support (RHEL) and added new DKMS drivers Updated product registration URL Updated standards and directives EC Conformity, pg. 7 Updated year of ICES-003 standard FCC & ICES-003 Compliance, pg. 7 Corrected TRIG IN Input type, removed: (5 V compliant) Auxiliary I/O input coupling (DC) Updated software descriptions and added order number for ATS-GPU-NUFFT Changes from version 1.5F (May 2019) to version 1.5G Changes from version 1.5F (May 2019) to version 1.5G Section, Page Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clock, replaced min. and max. amplitude with a note that signal levels specified on page 8 must be respected. Removed qualified metrology lab as option for recalibrating ATS9371 Specified Unindows 7 version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from *400 mVp_p" to *400 mVp_p to 2 Vp_p" Optional ECLK (External Clock) Input, pg. 8 Removed maximum amplitude, information is included in signal level Changed signal level from *200 mV to 400 mVp_p to 2 Vp_p" Optional 10 MHz Reference PLL Input, pg. 8 Removed ansimum amplitude, information added Pacer Output) Auxiliary I/O (AUX I/O), pg. 9 Changes from version 1.5E (Jan 2019) to version 1.5F Section, Page Changes from version as this product is being discontinued ATS-GPU, pg. 6 Removed ATS-GMA section as this product is being discontinued ATS-GMA, pg. 6 Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Ordering Information, pg. 9 Updated Trademark information	Updated OCT Ignore Bad Clock URL	OCT Ignore Bad Clock, pg. 5
Updated product registration URL Updated standards and directives EC Conformity, pg. 7 Updated year of ICES-003 standard FCC & ICES-003 Compliance, pg. 7 Corrected TRIG IN Input type, removed: (5 V compliant) TRIG IN (External Trigger) Input, pg. 9 Added Auxiliary I/O input coupling (DC) Auxiliary I/O input (DC) Auxiliary I/O input coupling (DC) Auxiliary I/O input (DC	Updated section ATS-GPU and added paragraph on ATS-GPU-NUFFT	ATS-GPU, pg. 6
Updated standards and directives Updated year of ICES-003 standard FCC & ICES-003 Compliance, pg. 7 Corrected TRIG IN Input type, removed: (5 V compliant) TRIG IN (External Trigger) Input, pg. 9 Added Auxiliary I/O input coupling (DC) Updated software descriptions and added order number for ATS-GPU-NUFFT Changes from version 1.5F (May 2019) to version 1.5G Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clock, replaced min. and max. amplitude with a note that signal levels specified on page 8 must be respected. Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mVp.p," to "400 mVp.p.p" to "400 mVp.p.p" to "400 mVp.p.p" Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mVp.p.p. to 200 mVp.p.p. (Optional ECLK (External Clock) Input, pg. 8 Removed of sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) Removed ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) Removed ATS-GMA section as this product is being discontinued ATS-GPU, pg. 6 Removed ATS-GMA section as this product is being discontinued ATS-GPU, pg. 6 Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Ordering Information, pg. 9 Updated Trademark information	Updated Linux Support (RHEL) and added new DKMS drivers	Linux Support, pg. 6
Updated year of ICES-003 standard Corrected TRIG IN Input type, removed: (5 V compliant) Added Auxiliary I/O input coupling (DC) Updated software descriptions and added order number for ATS-GPU-NUFFT Cordering Information, pg. 9 Changes from version 1.5F (May 2019) to version 1.5G Changed Sampling Rate column to Max. Sample Rate Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clock, replaced min. and max. amplitude with a note that signal levels specified on page 8 must be respected. Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P.p.p} " to Public Removed maximum amplitude, information is included in signal level Changed signal level from *200 mV to 400 mV _{P.p.p} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) Removed ATS-GMA section as this product is being discontinued ATS-GPU, pg. 6 Extended Warranty, pg. 6 Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Ordering Information, pg. 9 Updated Trademark information Section, Page Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page	Updated product registration URL	Extended Warranty, pg. 7
Corrected TRIG IN Input type, removed: (5 V compliant) Added Auxiliary I/O input coupling (DC) Updated software descriptions and added order number for ATS-GPU-NUFFT Changes from version 1.5F (May 2019) to version 1.5G Changed Sampling Rate column to Max. Sample Rate Changed Signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clock, replaced min. and max. amplitude with a note that signal levels specified on page 8 must be respected. Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{D-P} " to "400 mV _{D-P} " to 2 V _{D-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{D-P} to 2 V _{D-P} " Removed Signal level from ±200 mV to 400 mV _{D-P} to Torrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) Removed ATS-GMA section as this product is being discontinued Added section Extended Warranty Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Updated Trademark information Description 1.5E Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page Changes from version 1.5D (Jan 2019) to version 1.5E	Updated standards and directives	EC Conformity, pg. 7
Added Auxiliary I/O input coupling (DC) Updated software descriptions and added order number for ATS-GPU-NUFFT Changes from version 1.5F (May 2019) to version 1.5G Changes from version 1.5F (May 2019) to version 1.5G Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clock, replaced min. and max. amplitude with a note that signal levels specified on page 8 must be respected. Removed qualified metrology lab as option for recalibrating ATS9371 Calibration, pg. 5 Specified Windows 7 version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P.P.} " to "400 mV _{P.P.} to 2 V _{P.P.} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P.P.} to 2 V _{P.P.} " Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) ATS-GPU, pg. 6 Removed ATS-GMA section as this product is being discontinued ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Ordering Information, pg. 9 Updated Trademark information Changes from version 1.5D (Jan 2019) to version 1.5E Changes from version 1.5D (Jan 2019) to version 1.5E Changes from version 1.5D (Jan 2019) to version 1.5E Changes from version 1.5D (Jan 2019) to version 1.5E Changes from version 1.5D (Jan 2019) to version 1.5E Changes from version 1.5D (Jan 2019) to version 1.5E Changes from version 1.5D (Jan 2019) to version 1.5E	Updated year of ICES-003 standard	FCC & ICES-003 Compliance, pg. 7
Changes from version 1.5F (May 2019) to version 1.5G Changes from version 1.5F (May 2019) to version 1.5G Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clock, replaced min. and max. amplitude with a note that signal levels specified on page 8 must be respected. Removed qualified metrology lab as option for recalibrating ATS9371 Calibration, pg. 5 Specified Windows 7 version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed Sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) Removed ATS-GMA section as this product is being discontinued ATS-GMA, pg. 6 Added section Extended Warranty Extended Warranty Extended Warranty Grdering Information, pg. 9 Changes from version 1.5D (Jan 2019) to version 1.5E Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page	Corrected TRIG IN Input type, removed: (5 V compliant)	TRIG IN (External Trigger) Input, pg. 9
Changes from version 1.5F (May 2019) to version 1.5G Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clock, replaced min. and max. amplitude with a note that signal levels specified on page 8 must be respected. Removed qualified metrology lab as option for recalibrating ATS9371 Calibration, pg. 5 Specified Windows 7 version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPM section as this product is being discontinued ATS-GPM, pg. 6 Removed ATS-GMA section as this product is being discontinued ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Ordering Information, pg. 9 Updated Trademark information Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page	Added Auxiliary I/O input coupling (DC)	Auxiliary I/O (AUX I/O), pg. 9
Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clock, replaced min. and max. amplitude with a note that signal levels specified on page 8 must be respected. Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Linux Support, pg. 6 Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) ATS-GPU, pg. 6 Removed ATS-GMA section as this product is being discontinued ATS-GMA section as this product is being discontinued ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Ordering Information, pg. 9 Updated Trademark information Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page	Updated software descriptions and added order number for ATS-GPU-NUFFT	Ordering Information, pg. 9
Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clock, replaced min. and max. amplitude with a note that signal levels specified on page 8 must be respected. Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} " to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} to 2 V _{P-P} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) Removed ATS-GMA section as this product is being discontinued ATS-GMA section as this product is being discontinued ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Ordering Information, pg. 9 Updated Trademark information Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page		
Because signal levels differ for Fast External Clock and 10 MHz Reference Clock, replaced min. and max. amplitude with a note that signal levels specified on page 8 must be respected. Removed qualified metrology lab as option for recalibrating ATS9371 Calibration, pg. 5 Specified Windows 7 version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Linux Support, pg. 6 Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Optional ECLK (External Clock) Input, pg. 8 Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) ATS-GPU, pg. 6 Removed ATS-GMA section as this product is being discontinued ATS-GMA, pg. 6 Added section Extended Warranty Extended Warranty Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Ordering Information, pg. 9 Updated Trademark information Deg. 9 Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page	Changes from version 1.5F (May 2019) to version 1.5G	Section, Page
Specified Windows 7 version support, re-ordered list of operating systems, and added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) Removed ATS-GMA section as this product is being discontinued Added section Extended Warranty Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Updated Trademark information Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page Section, Page		
added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) ATS-GPU, pg. 6 Removed ATS-GMA section as this product is being discontinued Added section Extended Warranty Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Updated Trademark information Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page	Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clo	Feature Table, pg. 1 Optional External Clock, pg. 5 ock, replaced
Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) Removed ATS-GMA section as this product is being discontinued Added section Extended Warranty Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Updated Trademark information Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page Extended Warranty, pg. 6 Ordering Information, pg. 9 Updated Trademark information Section, Page	Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clo min. and max. amplitude with a note that signal levels specified on page 8 m	Feature Table, pg. 1 Optional External Clock, pg. 5 ock, replaced nust be respected.
Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) Removed ATS-GMA section as this product is being discontinued Added section Extended Warranty Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Updated Trademark information Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page Section, Page Extended Warranty, pg. 6 Ordering Information, pg. 9 Updated Trademark information Section, Page	Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clo min. and max. amplitude with a note that signal levels specified on page 8 m Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, an	Feature Table, pg. 1 Optional External Clock, pg. 5 ock, replaced nust be respected. Calibration, pg. 5
Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) Removed ATS-GMA section as this product is being discontinued Added section Extended Warranty Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Updated Trademark information Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page Section, Page	Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clo min. and max. amplitude with a note that signal levels specified on page 8 m Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, an added end-of-support notice for Windows 7 and Windows Server 2008 R2	Optional External Clock, pg. 5 ock, replaced nust be respected. Calibration, pg. 5 d Support for Windows, pg. 6
Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) Removed ATS-GMA section as this product is being discontinued Added section Extended Warranty Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Updated Trademark information Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page Section, Page	Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clo min. and max. amplitude with a note that signal levels specified on page 8 m Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, an added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} "	Optional External Clock, pg. 5 ock, replaced nust be respected. Calibration, pg. 5 d Support for Windows, pg. 6 Linux Support, pg. 6
Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GPU) Removed ATS-GMA section as this product is being discontinued ATS-GMA, pg. 6 Added section Extended Warranty Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Updated Trademark information Changes from version 1.5D (Jan 2019) to version 1.5E ATS-GPU, pg. 6 Extended Warranty, pg. 6 Ordering Information, pg. 9 pg. 9	Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clo min. and max. amplitude with a note that signal levels specified on page 8 m Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, an added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P}	Optional External Clock, pg. 1 Optional External Clock, pg. 5 ock, replaced nust be respected. Calibration, pg. 5 d Support for Windows, pg. 6 Linux Support, pg. 6 Optional ECLK (External Clock) Input, pg. 8
Removed ATS-GMA section as this product is being discontinued Added section Extended Warranty Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Updated Trademark information Changes from version 1.5D (Jan 2019) to version 1.5E ATS-GMA, pg. 6 Extended Warranty, pg. 6 Ordering Information, pg. 9 pg. 9	Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clo min. and max. amplitude with a note that signal levels specified on page 8 m Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, an added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed sine or square wave requirement	Optional External Clock, pg. 1 Optional External Clock, pg. 5 ock, replaced nust be respected. Calibration, pg. 5 d Support for Windows, pg. 6 Linux Support, pg. 6 Optional ECLK (External Clock) Input, pg. 8 Optional 10 MHz Reference PLL Input, pg. 8
Added section Extended Warranty Extended Warranty, pg. 6 Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Ordering Information, pg. 9 Updated Trademark information pg. 9 Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page	Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clo min. and max. amplitude with a note that signal levels specified on page 8 m Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, an added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output)	Feature Table, pg. 1 Optional External Clock, pg. 5 ock, replaced nust be respected. Calibration, pg. 5 dd Support for Windows, pg. 6 Linux Support, pg. 6 Optional ECLK (External Clock) Input, pg. 8 Optional 10 MHz Reference PLL Input, pg. 8 Auxiliary I/O (AUX I/O), pg. 9
Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Updated Trademark information Changes from version 1.5D (Jan 2019) to version 1.5E Ordering Information, pg. 9 pg. 9 Section, Page	Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clo min. and max. amplitude with a note that signal levels specified on page 8 m Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, an added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F	Optional External Clock, pg. 1 Optional External Clock, pg. 5 ock, replaced nust be respected. Calibration, pg. 5 Support for Windows, pg. 6 Linux Support, pg. 6 Optional ECLK (External Clock) Input, pg. 8 Optional 10 MHz Reference PLL Input, pg. 8 Auxiliary I/O (AUX I/O), pg. 9 Section, Page
Updated Trademark information pg. 9 Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page	Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clo min. and max. amplitude with a note that signal levels specified on page 8 m Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, an added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and Gi	Feature Table, pg. 1 Optional External Clock, pg. 5 ock, replaced oust be respected. Calibration, pg. 5 d Support for Windows, pg. 6 Linux Support, pg. 6 Optional ECLK (External Clock) Input, pg. 8 Optional 10 MHz Reference PLL Input, pg. 8 Auxiliary I/O (AUX I/O), pg. 9 Section, Page PU) ATS-GPU, pg. 6
Changes from version 1.5D (Jan 2019) to version 1.5E Section, Page	Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clo min. and max. amplitude with a note that signal levels specified on page 8 m Removed qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, an added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GI Removed ATS-GMA section as this product is being discontinued	Feature Table, pg. 1 Optional External Clock, pg. 5 ock, replaced nust be respected. Calibration, pg. 5 d Support for Windows, pg. 6 Linux Support, pg. 6 Optional ECLK (External Clock) Input, pg. 8 Optional 10 MHz Reference PLL Input, pg. 8 Auxiliary I/O (AUX I/O), pg. 9 Section, Page PU) ATS-GPU, pg. 6 ATS-GMA, pg. 6
	Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clomin. and max. amplitude with a note that signal levels specified on page 8 mRemoved qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, an added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and Gil Removed ATS-GMA section as this product is being discontinued Added section Extended Warranty	Feature Table, pg. 1 Optional External Clock, pg. 5 ock, replaced nust be respected. Calibration, pg. 5 d Support for Windows, pg. 6 Linux Support, pg. 6 Optional ECLK (External Clock) Input, pg. 8 Optional 10 MHz Reference PLL Input, pg. 8 Auxiliary I/O (AUX I/O), pg. 9 Section, Page PU) ATS-GPU, pg. 6 ATS-GMA, pg. 6 Extended Warranty, pg. 6
Updated Sanctions List URL Export Control Classification, pg. 6	Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clomin. and max. amplitude with a note that signal levels specified on page 8 mRemoved qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, an added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and Gl Removed ATS-GMA section as this product is being discontinued Added section Extended Warranty Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101)	PU) Feature Table, pg. 1 Optional External Clock, pg. 5 Cock, replaced Fock, replaced Calibration, pg. 5 Support for Windows, pg. 6 Linux Support, pg. 6 Optional ECLK (External Clock) Input, pg. 8 Optional 10 MHz Reference PLL Input, pg. 8 Auxiliary I/O (AUX I/O), pg. 9 Section, Page PU) ATS-GPU, pg. 6 ATS-GMA, pg. 6 Extended Warranty, pg. 6 Ordering Information, pg. 9
	Changed Sampling Rate column to Max. Sample Rate Replaced signal sine or square wave requirement with high slew rate Because signal levels differ for Fast External Clock and 10 MHz Reference Clomin. and max. amplitude with a note that signal levels specified on page 8 mRemoved qualified metrology lab as option for recalibrating ATS9371 Specified Windows 7 version support, re-ordered list of operating systems, an added end-of-support notice for Windows 7 and Windows Server 2008 R2 Specified Linux distributions: CentOS, Debian, and Ubuntu Changed signal level from "400 mV _{P-P} " to "400 mV _{P-P} to 2 V _{P-P} " Removed maximum amplitude, information is included in signal level Changed signal level from ±200 mV to 400 mV _{P-P} Removed sine or square wave requirement Corrected Output types (removed Busy Output and added Pacer Output) Changes from version 1.5E (Jan 2019) to version 1.5F Updated ATS-GPU data transfer rate and benchmarks (FFTs per second and GIR Removed ATS-GMA section as this product is being discontinued Added section Extended Warranty Removed ATS-GMA order numbers (ATSGMA-001, ATSGMA-101) Updated Trademark information	PU) Feature Table, pg. 1 Optional External Clock, pg. 5 Calibration, pg. 5 Support for Windows, pg. 6 Linux Support, pg. 6 Optional ECLK (External Clock) Input, pg. 8 Optional 10 MHz Reference PLL Input, pg. 8 Auxiliary I/O (AUX I/O), pg. 9 Section, Page PU) ATS-GPU, pg. 6 ATS-GMA, pg. 6 Extended Warranty, pg. 6 Ordering Information, pg. 9 pg. 9



DATASHEET REVISION HISTORY Changes from version 1.5C (Dec 2018) to version 1.5D Section, Page Corrected upper limit of Fast External Clock frequency to 1 GHz Fast External Clock, pg. 4 Corrected upper limit of sampling clock to 1 GHz 10 MHz Reference Clock, pg. 5 Corrected upper limit of external clock frequency to 1 GHz OCT Ignore Bad Clock, pg. 5 Changes from version 1.5B (Sept 2018) to version 1.5C Section, Page Added note about LSB being replaced by the state of the external trigger signal source External Trigger Input, pg. 4 Changes from version 1.5A (Jan 2018) to version 1.5B Section, Page Updated RoHS Compliance to RoHS 3 Global change Updated product image pg. 1 Clarified Operating System Support Feature Table, pg. 1 Updated Recommended Motherboards or PCs Recommended Motherboards or PCs, pg. 2 Correction of trigger engines: changed to J and K (instead of X and Y) Triggering, pg. 4 Specified that External Trigger Input 3.3 V TTL input is 5 V-compliant, and External Trigger Input, pg. 4 added note about LSB being replaced by the state of the external trigger signal source Added information on ATS-SDK license Software Development Kits, pg. 5 Specified 64-bit version for Windows and Linux support ATS-GPU, pg. 6 Added ATS-GMA section ATS-GMA, pg. 6 Added list of supported Microsoft Windows versions Support for Windows, pg. 6 Added Acquisition Memory System section Acquisition Memory System, pg. 8 Optional ECLK (External Clock) Input, pg. 8 Clarified that Max. and Min. Frequencies are for Fast External Clock, corrected Maximum Frequency, and added Maximum Amplitude: 2 V_{P-P} Added "PLL" to section name for clarity Optional 10 MHz Reference PLL Input, pg. 8 Corrected Input Frequency tolerance and added Max. and Min. Frequencies Corrected upper limit of Sampling Clock Frequency range to 1 GHz Corrected Trigger Engine Combination Triggering System, pg. 8 Clarified specs by providing separate specifications for Analog and TTL input, TRIG IN (External Trigger) Input, pg. 9 Added TTL min. pulse width, TTL min. pulse amplitude, and TTL input protection Replaced TRIG OUT Output section with Auxiliary I/O (AUX I/O), and Auxiliary I/O (AUX I/O), pg. 9 corrected Input Amplitude: 3.3 V TTL (5 V-compliant) Added subscription length for ATS-SDK, ATSGPU-001, ATSGPU-101 Ordering Information, pg. 9 Added products ATSGMA-001, ATSGMA-101

Changes from version 1.5 (Sept 2017) to version 1.5A

Added Trademark information

Added FFT length	Overview, pg. 1
Added note about NPT Footers	No Pre-Trigger (NPT) AutoDMA, pg. 3
Added note about NPT Footers not being available in Data Packing Mode	Data Packing Mode, pg. 4
Edited section on OCT Ignore Bad Clock	OCT Ignore Bad Clock, pg. 5
Added note about Trigger Enable Input use in OCT	AUX Connector, pg. 5
Added -BASE and -OCT to ATS-GPU description for clarity	ATS-GPU, pg. 5
Corrected size of card	Physical, pg. 7
Updated email address	Manufactured Bv. pg. 8

pg. 9

Section, Page